

CTS-10532

Seal Wash Face Seal
Equivalent to: Waters® WAT271017



Technical Data Sheet

1.0 Description

This seal is designed to be compatible with Waters 2690, 2690D, 2695, 2695D, 2790, 2795, 2796 and Alliance HPLC systems. It is equivalent to Waters part number WAT271017.

2.0 Wetted Materials Specifications

PTFE
Elgiloy Spring

3.0 Replacement Frequency

Beyond routine maintenance it may be necessary to replace the seals more frequently if they are operating in a high salt buffer concentration environment or if the plunger themselves are scratched and worn. Seals may also need replacing if the system fails its leak test.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.