

# CTS-A11187

Outlet Check Valve

Equivalent to: Shimadzu® 228-45705-91

Old# 228-45563-95



## Technical Data Sheet

---

### 1.0 Description

This check valve is designed to be compatible with Shimadzu i-Series, LC-10ADvp, LC-10ATvp, LC-20AD/AB, LC-20ADXR and LC-30ADSF pumps. It is equivalent to Shimadzu part number 228-45705-91. It replaces old part number 228-45563-95.

### 2.0 Materials Specifications

316 SS

PEEK™

Ruby

Sapphire

### 3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



### 4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.